

Specifications		Free and Total Chlorine HR
Measurement	Range (all methods)	0.00 to 10.00 mg/L (as Cl <sub>2</sub> )
	Resolution (all methods)	0.01 mg/L
	Accuracy @25°C (77°F) (all methods)	±0.03 mg/L ±3% of reading at 25 °C
	Method	Adaptation of the EPA DPD method 330.5
Measurement System	Light Source	light emitting diode
	Bandpass filter	525 nm
	Bandpass filter bandwidth	8 nm
	Bandpass filter wavelength accuracy	±1.0 nm
	Light Detector	silicon photocell
	Cuvette type	round 24.6 mm diameter (22 mm inside)
Additional Specifications	Auto logging	50 readings
	Display	128 x 64 pixel B/W LCD with backlight
	Auto-off	after 15 minutes of inactivity (30 minutes before a READ measurement)
	Battery type / Life	alkaline 1.5 V AA (3) / > 800 measurements (without backlight)

HI97734

# 

380 g (13.4 oz.)

#### Ordering Information

 $\label{eq:Higher_Higher_Higher_Higher} \textbf{Higher} \textbf{a} \textbf{Check standards}, \textbf{sample cuvettes (2)}, \textbf{sample caps (2)}, \textbf{plastic stoppers (2)}, 1.5 V AA batteries (3), cuvette wiping cloth, scissors, instrument quality certificate, instruction manual and HI7101412 rigid carrying case.$ 

142.5 x 102.5 x 50.5 mm (5.6 x 4.0 x 2.0")

HI97734 is supplied with sample cuvettes (2), sample caps (2), plastic stoppers (2), 1.5V AA

Reagents sold separately

Environment

Dimensions

Weight

# Reagents and Standards

HI97734

HI97734-11 CAL Check standard cuvettes for free and total chlorine high range

0 to 50°C (32 to 122°F); 0 to 100% RH, non-serviceable

HI93734-01 free and total chlorine HR reagent (100 tests)
HI93734-03 free and total chlorine HR reagent (300 tests)

#### HI97734

# Free and Total Chlorine High Range Portable Photometer

## • Advanced LED optical system

 LEDs have a much higher luminous efficiency, providing more light while using less power. They also produce little heat, which could otherwise affect electronic stability.

#### CAL Check™

 Validate the performance of the instrument at any time and apply a user calibration using CAL Check cuvettes made with NIST traceable standards. The CAL Check screen guides the user step-by-step through the validation process and user calibration

#### Auto logging

- GLP
  - Review of the last calibration date.

#### • On-screen tutorial mode with animations

- The built in Tutorial mode guides users step-by-step through the measurement process.
- Waterproof and floating IP67 case
- Auto-shut off
  - Automatic shut off after 15 minutes of inactivity and 30 minutes before a read measurement.

#### • Battery status indicator

- Indicates the amount of battery life left.
- · Built-in timer
  - Display of time remaining before a measurement is taken. Ensures that all readings are taken at the appropriate reaction intervals for the test being performed.

### • Error messages

 Messages on display alerting to problems including no cap, high zero, and standard too low.

## Units of measure

 Appropriate unit of measure is displayed along with reading.

